(2012 OSHA Hazard Communication Standard (29 CFR 1910.1200))

SAFETY DATA SHEET

SIEMENS

Atellica™ IM Hepatitis B surface Antigen II (HBsII)

SDS # : 10995604

Section 1. Identification

Product identifier : Atellica™ IM Hepatitis B surface Antigen II (HBsII)

Product code : 10995604

Product type : Liquid.

Relevant identified uses of the substance or mixture and uses advised against

Not applicable.

Manufactured/supplied : Siemens Healthcare Diagnostics Inc.
511 Benedict Avenue
Tarrytown, NY 10591-5097 USA
1-877-229-3711

(800) 424-9300 (CHEMTREC) (24/365)

Section 2. Hazards identification

OSHA/HCS status :

HBsII Lite Reagent	While this material is not considered hazardous by the OSHA Hazard Communication Standard (29 CFR 1910. 1200), this SDS contains valuable information critical to the safe handling and proper use of the product. This SDS should be retained and available for employees and other users of this product.
HBsII Solid Phase Reagent	While this material is not considered hazardous by the OSHA Hazard Communication Standard (29 CFR 1910. 1200), this SDS contains valuable information critical to the safe handling and proper use of the product. This SDS should be retained and available for employees and other users of this product.
HBsII Ancillary Reagent	While this material is not considered hazardous by the OSHA Hazard Communication Standard (29 CFR 1910. 1200), this SDS contains valuable information critical to the safe handling and proper use of the product. This SDS should be retained and available for employees and other users of this product.
HBsII Low Calibrator	While this material is not considered hazardous by the OSHA Hazard Communication Standard (29 CFR 1910. 1200), this SDS contains valuable information critical to the safe handling and proper use of the product. This SDS should be retained and available for employees and other users of this product.
HBsII High Calibrator	While this material is not considered hazardous by the OSHA Hazard Communication Standard (29 CFR 1910. 1200), this SDS contains valuable information critical to the safe handling and proper use of the product. This SDS should be retained and available for employees

Section 2. Hazards identification

and other users of this product.

Classification of the substance or mixture :

HBsII Lite Reagent	Not classified.
HBsII Solid Phase Reagent	Not classified.
HBsII Ancillary Reagent	Not classified.
HBsII Low Calibrator	Not classified.
HBsII High Calibrator	Not classified.

Additional information : Potentially biohazardous material.

Sodium azide may react with lead or copper plumbing to form highly explosive metal azides.

GHS label elements

Signal word :

HBsII Lite Reagent	No signal word.
HBsII Solid Phase Reagent	No signal word.
HBsII Ancillary Reagent	No signal word.
HBsII Low Calibrator	No signal word.
HBsII High Calibrator	No signal word.

Hazard statements :

HBsII Lite Reagent	No known significant effects or critical hazards.
HBsII Solid Phase Reagent	No known significant effects or critical hazards.
HBsII Ancillary Reagent	No known significant effects or critical hazards.
HBsII Low Calibrator	No known significant effects or critical hazards.
HBsII High Calibrator	No known significant effects or critical hazards.

Precautionary statements

Prevention :

HBsII Lite Reagent	Not applicable.
HBsII Solid Phase Reagent	Not applicable.
HBsII Ancillary Reagent	Not applicable.
HBsII Low Calibrator	Not applicable.
HBsII High Calibrator	Not applicable.

Response :

HBsII Lite Reagent	Not applicable.
HBsII Solid Phase Reagent	Not applicable.
HBsII Ancillary Reagent	Not applicable.
HBsII Low Calibrator	Not applicable.
HBsII High Calibrator	Not applicable.

Storage :

HBsII Lite Reagent	Not applicable.
HBsII Solid Phase Reagent	Not applicable.
HBsII Ancillary Reagent	Not applicable.
HBsII Low Calibrator	Not applicable.
HBsII High Calibrator	Not applicable.

Disposal :

HBsII Lite Reagent	Not applicable.
HBsII Solid Phase Reagent	Not applicable.
HBsII Ancillary Reagent	Not applicable.
HBsII Low Calibrator	Not applicable.
HBsII High Calibrator	Not applicable.

Supplemental label elements :

HBsII Lite Reagent	None known.
HBsII Solid Phase Reagent	None known.
HBsII Ancillary Reagent	None known.
HBsII Low Calibrator	None known.
HBsII High Calibrator	None known.

Hazards not otherwise classified :

HBsII Lite Reagent	None known.
HBsII Solid Phase Reagent	None known.
HBsII Ancillary Reagent	None known.
HBsII Low Calibrator	None known.
HBsII High Calibrator	None known.

Section 3. Composition/information on ingredients

Substance/mixture : HBsII Lite Reagent — Mixture
HBsII Solid Phase Reagent — Mixture
HBsII Ancillary Reagent — Mixture
HBsII Low Calibrator — Mixture
HBsII High Calibrator — Mixture

Ingredient name	%	CAS number
HBsII Lite Reagent		
sodium azide	0.09	26628-22-8
HBsII Solid Phase Reagent		
sodium azide	0.09	26628-22-8
HBsII Ancillary Reagent		
sodium azide	0.09	26628-22-8
HBsII Low Calibrator		
sodium azide	0.09	26628-22-8
HBsII High Calibrator		
sodium azide	0.09	26628-22-8

Any concentration shown as a range is to protect confidentiality or is due to batch variation.

There are no additional ingredients present which, within the current knowledge of the supplier and in the concentrations applicable, are classified as hazardous to health or the environment and hence require reporting in this section.

Occupational exposure limits, if available, are listed in Section 8.

Section 4. First aid measures

Description of necessary first aid measures

Eye contact : HBsII Lite Reagent — Immediately flush eyes with plenty of water, occasionally lifting the upper and lower eyelids. Check for and remove any contact lenses. Get medical attention if irritation occurs.

HBsII Solid Phase Reagent — Immediately flush eyes with plenty of water, occasionally lifting the upper and lower eyelids. Check for and remove any contact lenses. Get medical attention if irritation occurs.

HBsII Ancillary Reagent — Immediately flush eyes with plenty of water, occasionally lifting the upper and lower eyelids. Check for and remove any contact lenses. Get medical attention if irritation occurs.

HBsII Low Calibrator — Immediately flush eyes with plenty of water, occasionally lifting the upper and lower eyelids. Check for and remove any contact lenses. Get medical attention if irritation occurs.

HBsII High Calibrator — Immediately flush eyes with plenty of water, occasionally lifting the upper and lower eyelids. Check for and remove any contact lenses. Get medical attention if irritation occurs.

Section 4. First aid measures

Inhalation	:	HBsII Lite Reagent	Remove victim to fresh air and keep at rest in a position comfortable for breathing. Get medical attention if symptoms occur.
		HBsII Solid Phase Reagent	Remove victim to fresh air and keep at rest in a position comfortable for breathing. Get medical attention if symptoms occur.
		HBsII Ancillary Reagent	Remove victim to fresh air and keep at rest in a position comfortable for breathing. Get medical attention if symptoms occur.
		HBsII Low Calibrator	Remove victim to fresh air and keep at rest in a position comfortable for breathing. Get medical attention if symptoms occur.
		HBsII High Calibrator	Remove victim to fresh air and keep at rest in a position comfortable for breathing. Get medical attention if symptoms occur.
Skin contact	:	HBsII Lite Reagent	Flush contaminated skin with plenty of water. Remove contaminated clothing and shoes. Get medical attention if symptoms occur.
		HBsII Solid Phase Reagent	Flush contaminated skin with plenty of water. Remove contaminated clothing and shoes. Get medical attention if symptoms occur.
		HBsII Ancillary Reagent	Flush contaminated skin with plenty of water. Remove contaminated clothing and shoes. Get medical attention if symptoms occur.
		HBsII Low Calibrator	Flush contaminated skin with plenty of water. Remove contaminated clothing and shoes. Get medical attention if symptoms occur.
		HBsII High Calibrator	Flush contaminated skin with plenty of water. Remove contaminated clothing and shoes. Get medical attention if symptoms occur.
Ingestion	:	HBsII Lite Reagent	Wash out mouth with water. Remove victim to fresh air and keep at rest in a position comfortable for breathing. If material has been swallowed and the exposed person is conscious, give small quantities of water to drink. Do not induce vomiting unless directed to do so by medical personnel. Get medical attention if symptoms occur.
		HBsII Solid Phase Reagent	Wash out mouth with water. Remove victim to fresh air and keep at rest in a position comfortable for breathing. If material has been swallowed and the exposed person is conscious, give small quantities of water to drink. Do not induce vomiting unless directed to do so by medical personnel. Get medical attention if symptoms occur.
		HBsII Ancillary Reagent	Wash out mouth with water. Remove victim to fresh air and keep at rest in a position comfortable for breathing. If material has been swallowed and the exposed person is conscious, give small quantities of water to drink. Do not induce vomiting unless directed to do so by medical personnel. Get medical attention if symptoms occur.
		HBsII Low Calibrator	Wash out mouth with water. Remove victim to fresh air and keep at rest in a

Section 4. First aid measures

		position comfortable for breathing. If material has been swallowed and the exposed person is conscious, give small quantities of water to drink. Do not induce vomiting unless directed to do so by medical personnel. Get medical attention if symptoms occur.
	HBsII High Calibrator	Wash out mouth with water. Remove victim to fresh air and keep at rest in a position comfortable for breathing. If material has been swallowed and the exposed person is conscious, give small quantities of water to drink. Do not induce vomiting unless directed to do so by medical personnel. Get medical attention if symptoms occur.

Most important symptoms/effects, acute and delayed

Potential acute health effects

Eye contact	:	HBsII Lite Reagent	No known significant effects or critical hazards.
		HBsII Solid Phase Reagent	No known significant effects or critical hazards.
		HBsII Ancillary Reagent	No known significant effects or critical hazards.
		HBsII Low Calibrator	No known significant effects or critical hazards.
		HBsII High Calibrator	No known significant effects or critical hazards.
Inhalation	:	HBsII Lite Reagent	No known significant effects or critical hazards.
		HBsII Solid Phase Reagent	No known significant effects or critical hazards.
		HBsII Ancillary Reagent	No known significant effects or critical hazards.
		HBsII Low Calibrator	No known significant effects or critical hazards.
		HBsII High Calibrator	No known significant effects or critical hazards.
Skin contact	:	HBsII Lite Reagent	No known significant effects or critical hazards.
		HBsII Solid Phase Reagent	No known significant effects or critical hazards.
		HBsII Ancillary Reagent	No known significant effects or critical hazards.
		HBsII Low Calibrator	No known significant effects or critical hazards.
		HBsII High Calibrator	No known significant effects or critical hazards.
Ingestion	:	HBsII Lite Reagent	No known significant effects or critical hazards.
		HBsII Solid Phase Reagent	No known significant effects or critical hazards.
		HBsII Ancillary Reagent	No known significant effects or critical hazards.
		HBsII Low Calibrator	No known significant effects or critical hazards.
		HBsII High Calibrator	No known significant effects or critical hazards.

Over-exposure signs/symptoms

Section 4. First aid measures

Eye contact	:	HBsII Lite Reagent	No specific data.
		HBsII Solid Phase Reagent	No specific data.
		HBsII Ancillary Reagent	No specific data.
		HBsII Low Calibrator	No specific data.
		HBsII High Calibrator	No specific data.
Inhalation	:	HBsII Lite Reagent	No specific data.
		HBsII Solid Phase Reagent	No specific data.
		HBsII Ancillary Reagent	No specific data.
		HBsII Low Calibrator	No specific data.
		HBsII High Calibrator	No specific data.
Skin contact	:	HBsII Lite Reagent	No specific data.
		HBsII Solid Phase Reagent	No specific data.
		HBsII Ancillary Reagent	No specific data.
		HBsII Low Calibrator	No specific data.
		HBsII High Calibrator	No specific data.
Ingestion	:	HBsII Lite Reagent	No specific data.
		HBsII Solid Phase Reagent	No specific data.
		HBsII Ancillary Reagent	No specific data.
		HBsII Low Calibrator	No specific data.
		HBsII High Calibrator	No specific data.

Indication of immediate medical attention and special treatment needed, if necessary

Notes to physician : Treat symptomatically. Contact poison treatment specialist immediately if large quantities have been ingested or inhaled.

Specific treatments : No specific treatment.

Protection of first-aiders : No action shall be taken involving any personal risk or without suitable training.

See toxicological information (Section 11)

Section 5. Fire-fighting measures

Extinguishing media

Suitable extinguishing media : In case of fire, use water spray (fog), foam or dry chemical.

Unsuitable extinguishing media : None known.

Specific hazards arising from the chemical : In a fire or if heated, a pressure increase will occur and the container may burst.

Hazardous thermal decomposition products : Decomposition products may include the following materials:
halogenated compounds
metal oxide/oxides

Special protective actions for fire-fighters : Promptly isolate the scene by removing all persons from the vicinity of the incident if there is a fire. No action shall be taken involving any personal risk or without suitable training.

Special protective equipment for fire-fighters : Fire-fighters should wear appropriate protective equipment and self-contained breathing apparatus (SCBA) with a full face-piece operated in positive pressure mode.

Section 6. Accidental release measures

Personal precautions, protective equipment and emergency procedures

For non-emergency personnel : No action shall be taken involving any personal risk or without suitable training. Evacuate surrounding areas. Keep unnecessary and unprotected personnel from entering. Do not touch or walk through spilled material. Put on appropriate personal protective equipment.

Section 6. Accidental release measures

For emergency responders : If specialized clothing is required to deal with the spillage, take note of any information in Section 8 on suitable and unsuitable materials. See also the information in "For non-emergency personnel".

Environmental precautions : Avoid dispersal of spilled material and runoff and contact with soil, waterways, drains and sewers. Inform the relevant authorities if the product has caused environmental pollution (sewers, waterways, soil or air).

Methods and materials for containment and cleaning up

Small spill : Stop leak if without risk. Move containers from spill area. Dilute with water and mop up if water-soluble. Alternatively, or if water-insoluble, absorb with an inert dry material and place in an appropriate waste disposal container. Dispose of via a licensed waste disposal contractor.

Large spill : Stop leak if without risk. Move containers from spill area. Prevent entry into sewers, water courses, basements or confined areas. Wash spillages into an effluent treatment plant or proceed as follows. Contain and collect spillage with non-combustible, absorbent material e.g. sand, earth, vermiculite or diatomaceous earth and place in container for disposal according to local regulations (see Section 13). Dispose of via a licensed waste disposal contractor. Note: see Section 1 for emergency contact information and Section 13 for waste disposal.

Section 7. Handling and storage

Precautions for safe handling

Protective measures : Put on appropriate personal protective equipment (see Section 8).

Advice on general occupational hygiene : Eating, drinking and smoking should be prohibited in areas where this material is handled, stored and processed. Workers should wash hands and face before eating, drinking and smoking. Remove contaminated clothing and protective equipment before entering eating areas. See also Section 8 for additional information on hygiene measures.

Conditions for safe storage, including any incompatibilities : Store in accordance with local regulations. Store in original container protected from direct sunlight in a dry, cool and well-ventilated area, away from incompatible materials (see Section 10) and food and drink. Keep container tightly closed and sealed until ready for use. Containers that have been opened must be carefully resealed and kept upright to prevent leakage. Do not store in unlabeled containers. Use appropriate containment to avoid environmental contamination.

Section 8. Exposure controls/personal protection

Control parameters

Occupational exposure limits

Ingredient name	Exposure limits
HBsII Lite Reagent sodium azide	**ACGIH TLV (United States, 3/2015). Notes: as hydrazoic acid vapor** C: 0.11 ppm, (as Hydrazoic acid vapor) Form: as Hydrazoic acid vapor **ACGIH TLV (United States, 3/2015).** C: 0.29 mg/m³, (as Sodium azide) Form: as Sodium azide **NIOSH REL (United States, 10/2013). Absorbed through skin. Notes: NAN3** CEIL: 0.3 mg/m³, (NAN3) **NIOSH REL (United States, 10/2013). Absorbed through skin. Notes: as HN3** CEIL: 0.1 ppm, (as HN3) **OSHA PEL 1989 (United States, 3/1989). Absorbed through skin. Notes: as HN3** CEIL: 0.1 ppm, (as HN3) **OSHA PEL 1989 (United States, 3/1989).**

Section 8. Exposure controls/personal protection

	Absorbed through skin. Notes: as NaN3 CEIL: 0.3 mg/m³, (as NaN3)
HBsII Solid Phase Reagent sodium azide	**ACGIH TLV (United States, 3/2015). Notes: as hydrazoic acid vapor** C: 0.11 ppm, (as Hydrazoic acid vapor) Form: as Hydrazoic acid vapor **ACGIH TLV (United States, 3/2015).** C: 0.29 mg/m³, (as Sodium azide) Form: as Sodium azide **NIOSH REL (United States, 10/2013). Absorbed through skin. Notes: NAN3** CEIL: 0.3 mg/m³, (NAN3) **NIOSH REL (United States, 10/2013). Absorbed through skin. Notes: as HN3** CEIL: 0.1 ppm, (as HN3) **OSHA PEL 1989 (United States, 3/1989). Absorbed through skin. Notes: as HN3** CEIL: 0.1 ppm, (as HN3) **OSHA PEL 1989 (United States, 3/1989). Absorbed through skin. Notes: as NaN3** CEIL: 0.3 mg/m³, (as NaN3)
HBsII Ancillary Reagent sodium azide	**ACGIH TLV (United States, 3/2015). Notes: as hydrazoic acid vapor** C: 0.11 ppm, (as Hydrazoic acid vapor) Form: as Hydrazoic acid vapor **ACGIH TLV (United States, 3/2015).** C: 0.29 mg/m³, (as Sodium azide) Form: as Sodium azide **NIOSH REL (United States, 10/2013). Absorbed through skin. Notes: NAN3** CEIL: 0.3 mg/m³, (NAN3) **NIOSH REL (United States, 10/2013). Absorbed through skin. Notes: as HN3** CEIL: 0.1 ppm, (as HN3) **OSHA PEL 1989 (United States, 3/1989). Absorbed through skin. Notes: as HN3** CEIL: 0.1 ppm, (as HN3) **OSHA PEL 1989 (United States, 3/1989). Absorbed through skin. Notes: as NaN3** CEIL: 0.3 mg/m³, (as NaN3)
HBsII Low Calibrator sodium azide	**ACGIH TLV (United States, 3/2015). Notes: as hydrazoic acid vapor** C: 0.11 ppm, (as Hydrazoic acid vapor) Form: as Hydrazoic acid vapor **ACGIH TLV (United States, 3/2015).** C: 0.29 mg/m³, (as Sodium azide) Form: as Sodium azide **NIOSH REL (United States, 10/2013). Absorbed through skin. Notes: NAN3** CEIL: 0.3 mg/m³, (NAN3) **NIOSH REL (United States, 10/2013). Absorbed through skin. Notes: as HN3** CEIL: 0.1 ppm, (as HN3) **OSHA PEL 1989 (United States, 3/1989). Absorbed through skin. Notes: as HN3** CEIL: 0.1 ppm, (as HN3)

Section 8. Exposure controls/personal protection

	OSHA PEL 1989 (United States, 3/1989). Absorbed through skin. Notes: as NaN3 CEIL: 0.3 mg/m³, (as NaN3)
HBsII High Calibrator sodium azide	**ACGIH TLV (United States, 3/2015). Notes: as hydrazoic acid vapor** C: 0.11 ppm, (as Hydrazoic acid vapor) Form: as Hydrazoic acid vapor **ACGIH TLV (United States, 3/2015).** C: 0.29 mg/m³, (as Sodium azide) Form: as Sodium azide **NIOSH REL (United States, 10/2013). Absorbed through skin. Notes: NAN3** CEIL: 0.3 mg/m³, (NAN3) **NIOSH REL (United States, 10/2013). Absorbed through skin. Notes: as HN3** CEIL: 0.1 ppm, (as HN3) **OSHA PEL 1989 (United States, 3/1989). Absorbed through skin. Notes: as HN3** CEIL: 0.1 ppm, (as HN3) **OSHA PEL 1989 (United States, 3/1989). Absorbed through skin. Notes: as NaN3** CEIL: 0.3 mg/m³, (as NaN3)

Appropriate engineering controls : Good general ventilation should be sufficient to control worker exposure to airborne contaminants.

Environmental exposure controls : Emissions from ventilation or work process equipment should be checked to ensure they comply with the requirements of environmental protection legislation. In some cases, fume scrubbers, filters or engineering modifications to the process equipment will be necessary to reduce emissions to acceptable levels.

Individual protection measures

Hygiene measures : Wash hands, forearms and face thoroughly after handling chemical products, before eating, smoking and using the lavatory and at the end of the working period. Appropriate techniques should be used to remove potentially contaminated clothing. Wash contaminated clothing before reusing. Ensure that eyewash stations and safety showers are close to the workstation location.

Eye/face protection : Safety eyewear complying with an approved standard should be used when a risk assessment indicates this is necessary to avoid exposure to liquid splashes, mists, gases or dusts. If contact is possible, the following protection should be worn, unless the assessment indicates a higher degree of protection: safety glasses with side-shields.

Skin protection

Hand protection : Chemical-resistant, impervious gloves complying with an approved standard should be worn at all times when handling chemical products if a risk assessment indicates this is necessary.

Body protection : Personal protective equipment for the body should be selected based on the task being performed and the risks involved and should be approved by a specialist before handling this product.

Other skin protection : Appropriate footwear and any additional skin protection measures should be selected based on the task being performed and the risks involved and should be approved by a specialist before handling this product.

Respiratory protection : Based on the hazard and potential for exposure, select a respirator that meets the appropriate standard or certification. Respirators must be used according to a respiratory protection program to ensure proper fitting, training, and other important aspects of use.

Section 9. Physical and chemical properties

Physical state	:	HBsII Lite Reagent	Liquid.
		HBsII Solid Phase Reagent	Liquid.
		HBsII Ancillary Reagent	Liquid.
		HBsII Low Calibrator	Liquid.
		HBsII High Calibrator	Liquid.
Color	:	HBsII Lite Reagent	Yellowish-brown.
		HBsII Solid Phase Reagent	Brown.
		HBsII Ancillary Reagent	Yellowish-brown.
		HBsII Low Calibrator	Colorless.
		HBsII High Calibrator	Colorless.
Odor	:	HBsII Lite Reagent	Odorless.
		HBsII Solid Phase Reagent	Odorless.
		HBsII Ancillary Reagent	Odorless.
		HBsII Low Calibrator	Odorless.
		HBsII High Calibrator	Odorless.
pH	:	HBsII Lite Reagent	7.1
		HBsII Solid Phase Reagent	7.05 to 7.15
		HBsII Ancillary Reagent	7.1
		HBsII Low Calibrator	6.5
		HBsII High Calibrator	6.5
Flash point	:	HBsII Lite Reagent	[Product does not sustain combustion.]
		HBsII Solid Phase Reagent	[Product does not sustain combustion.]
		HBsII Ancillary Reagent	[Product does not sustain combustion.]
		HBsII Low Calibrator	Not available.
		HBsII High Calibrator	Not available.
Flammability (solid, gas)	:	HBsII Lite Reagent	Not relevant/applicable due to nature of the product.
		HBsII Solid Phase Reagent	Not relevant/applicable due to nature of the product.
		HBsII Ancillary Reagent	Not relevant/applicable due to nature of the product.
		HBsII Low Calibrator	Not relevant/applicable due to nature of the product.
		HBsII High Calibrator	Not relevant/applicable due to nature of the product.
Relative density	:	HBsII Lite Reagent	Not relevant/applicable due to nature of the product.
		HBsII Solid Phase Reagent	Not relevant/applicable due to nature of the product.
		HBsII Ancillary Reagent	Not relevant/applicable due to nature of the product.
		HBsII Low Calibrator	1
		HBsII High Calibrator	1
Solubility in water	:	HBsII Lite Reagent	Not relevant/applicable due to nature of the product.
		HBsII Solid Phase Reagent	Not relevant/applicable due to nature of the product.
		HBsII Ancillary Reagent	Not relevant/applicable due to nature of the product.
		HBsII Low Calibrator	Not relevant/applicable due to nature of the product.
		HBsII High Calibrator	Not relevant/applicable due to nature of the product.
Partition coefficient: n-octanol/water	:	HBsII Lite Reagent	Not available.
		HBsII Solid Phase Reagent	Not available.
		HBsII Ancillary Reagent	Not available.
		HBsII Low Calibrator	Not available.
		HBsII High Calibrator	Not available.

Section 9. Physical and chemical properties

Auto-ignition temperature :

HBsII Lite Reagent	Not available.
HBsII Solid Phase Reagent	Not available.
HBsII Ancillary Reagent	Not available.
HBsII Low Calibrator	Not available.
HBsII High Calibrator	Not available.

Viscosity :

HBsII Lite Reagent	Not available.
HBsII Solid Phase Reagent	Not available.
HBsII Ancillary Reagent	Not available.
HBsII Low Calibrator	Not available.
HBsII High Calibrator	Not available.

Aerosol product

Type of aerosol :

HBsII Lite Reagent	Not applicable.
HBsII Solid Phase Reagent	Not applicable.
HBsII Ancillary Reagent	Not applicable.
HBsII Low Calibrator	Not applicable.
HBsII High Calibrator	Not applicable.

Section 10. Stability and reactivity

Reactivity :

HBsII Lite Reagent	No specific test data related to reactivity available for this product or its ingredients.
HBsII Solid Phase Reagent	No specific test data related to reactivity available for this product or its ingredients.
HBsII Ancillary Reagent	No specific test data related to reactivity available for this product or its ingredients.
HBsII Low Calibrator	No specific test data related to reactivity available for this product or its ingredients.
HBsII High Calibrator	No specific test data related to reactivity available for this product or its ingredients.

Chemical stability :

HBsII Lite Reagent	The product is stable.
HBsII Solid Phase Reagent	The product is stable.
HBsII Ancillary Reagent	The product is stable.
HBsII Low Calibrator	The product is stable.
HBsII High Calibrator	The product is stable.

Possibility of hazardous reactions : Under normal conditions of storage and use, hazardous reactions will not occur.

Conditions to avoid : No specific data.

Incompatible materials : No specific data.

Hazardous decomposition products : Under normal conditions of storage and use, hazardous decomposition products should not be produced.

Section 11. Toxicological information

Information on toxicological effects

Acute toxicity

Product/ingredient name	Result	Species	Dose	Exposure
HBsII Lite Reagent				
sodium azide	LD50 Dermal	Rabbit	20 mg/kg	-
	LD50 Dermal	Rat	50 mg/kg	-
	LD50 Oral	Rat	27 mg/kg	-
HBsII Solid Phase Reagent				
sodium azide	LD50 Dermal	Rabbit	20 mg/kg	-
	LD50 Dermal	Rat	50 mg/kg	-

Section 11. Toxicological information

	LD50 Oral	Rat	27 mg/kg	-
HBsII Ancillary Reagent				
sodium azide	LD50 Dermal	Rabbit	20 mg/kg	-
	LD50 Dermal	Rat	50 mg/kg	-
	LD50 Oral	Rat	27 mg/kg	-
HBsII Low Calibrator				
sodium azide	LD50 Dermal	Rabbit	20 mg/kg	-
	LD50 Dermal	Rat	50 mg/kg	-
	LD50 Oral	Rat	27 mg/kg	-
HBsII High Calibrator				
sodium azide	LD50 Dermal	Rabbit	20 mg/kg	-
	LD50 Dermal	Rat	50 mg/kg	-
	LD50 Oral	Rat	27 mg/kg	-

Conclusion/Summary :
HBsII Lite Reagent — Not available.
HBsII Solid Phase Reagent — Not available.
HBsII Ancillary Reagent — Not available.
HBsII Low Calibrator — Not available.
HBsII High Calibrator — Not available.

Irritation/Corrosion

Not available.

Conclusion/Summary

Skin :
HBsII Lite Reagent — Not available.
HBsII Solid Phase Reagent — Not available.
HBsII Ancillary Reagent — Not available.
HBsII Low Calibrator — Not available.
HBsII High Calibrator — Not available.

Eyes :
HBsII Lite Reagent — Not available.
HBsII Solid Phase Reagent — Not available.
HBsII Ancillary Reagent — Not available.
HBsII Low Calibrator — Not available.
HBsII High Calibrator — Not available.

Respiratory :
HBsII Lite Reagent — Not available.
HBsII Solid Phase Reagent — Not available.
HBsII Ancillary Reagent — Not available.
HBsII Low Calibrator — Not available.
HBsII High Calibrator — Not available.

Sensitization

Not available.

Conclusion/Summary

Skin :
HBsII Lite Reagent — Not available.
HBsII Solid Phase Reagent — Not available.
HBsII Ancillary Reagent — Not available.
HBsII Low Calibrator — Not available.
HBsII High Calibrator — Not available.

Respiratory :
HBsII Lite Reagent — Not available.
HBsII Solid Phase Reagent — Not available.
HBsII Ancillary Reagent — Not available.
HBsII Low Calibrator — Not available.
HBsII High Calibrator — Not available.

Mutagenicity

Not available.

Section 11. Toxicological information

Conclusion/Summary	:	HBsII Lite Reagent	Not available.
		HBsII Solid Phase Reagent	Not available.
		HBsII Ancillary Reagent	Not available.
		HBsII Low Calibrator	Not available.
		HBsII High Calibrator	Not available.

Carcinogenicity

Not available.

Conclusion/Summary	:	HBsII Lite Reagent	Not available.
		HBsII Solid Phase Reagent	Not available.
		HBsII Ancillary Reagent	Not available.
		HBsII Low Calibrator	Not available.
		HBsII High Calibrator	Not available.

Reproductive toxicity

Not available.

Conclusion/Summary	:	HBsII Lite Reagent	Not available.
		HBsII Solid Phase Reagent	Not available.
		HBsII Ancillary Reagent	Not available.
		HBsII Low Calibrator	Not available.
		HBsII High Calibrator	Not available.

Teratogenicity

Not available.

Conclusion/Summary	:	HBsII Lite Reagent	Not available.
		HBsII Solid Phase Reagent	Not available.
		HBsII Ancillary Reagent	Not available.
		HBsII Low Calibrator	Not available.
		HBsII High Calibrator	Not available.

Specific target organ toxicity (single exposure)

Not available.

Specific target organ toxicity (repeated exposure)

Not available.

Aspiration hazard

Not available.

Information on the likely routes of exposure	:	HBsII Lite Reagent	Not available.
		HBsII Solid Phase Reagent	Not available.
		HBsII Ancillary Reagent	Not available.
		HBsII Low Calibrator	Not available.
		HBsII High Calibrator	Not available.

Potential acute health effects

Eye contact	:	HBsII Lite Reagent	No known significant effects or critical hazards.
		HBsII Solid Phase Reagent	No known significant effects or critical hazards.
		HBsII Ancillary Reagent	No known significant effects or critical hazards.
		HBsII Low Calibrator	No known significant effects or critical hazards.
		HBsII High Calibrator	No known significant effects or critical hazards.
Inhalation	:	HBsII Lite Reagent	No known significant effects or critical hazards.
		HBsII Solid Phase Reagent	No known significant effects or critical hazards.
		HBsII Ancillary Reagent	No known significant effects or critical hazards.
		HBsII Low Calibrator	No known significant effects or critical

Section 11. Toxicological information

			hazards.
		HBsII High Calibrator	No known significant effects or critical hazards.
Skin contact	:	HBsII Lite Reagent	No known significant effects or critical hazards.
		HBsII Solid Phase Reagent	No known significant effects or critical hazards.
		HBsII Ancillary Reagent	No known significant effects or critical hazards.
		HBsII Low Calibrator	No known significant effects or critical hazards.
		HBsII High Calibrator	No known significant effects or critical hazards.
Ingestion	:	HBsII Lite Reagent	No known significant effects or critical hazards.
		HBsII Solid Phase Reagent	No known significant effects or critical hazards.
		HBsII Ancillary Reagent	No known significant effects or critical hazards.
		HBsII Low Calibrator	No known significant effects or critical hazards.
		HBsII High Calibrator	No known significant effects or critical hazards.

Symptoms related to the physical, chemical and toxicological characteristics

Eye contact	:	HBsII Lite Reagent	No specific data.
		HBsII Solid Phase Reagent	No specific data.
		HBsII Ancillary Reagent	No specific data.
		HBsII Low Calibrator	No specific data.
		HBsII High Calibrator	No specific data.
Inhalation	:	HBsII Lite Reagent	No specific data.
		HBsII Solid Phase Reagent	No specific data.
		HBsII Ancillary Reagent	No specific data.
		HBsII Low Calibrator	No specific data.
		HBsII High Calibrator	No specific data.
Skin contact	:	HBsII Lite Reagent	No specific data.
		HBsII Solid Phase Reagent	No specific data.
		HBsII Ancillary Reagent	No specific data.
		HBsII Low Calibrator	No specific data.
		HBsII High Calibrator	No specific data.
Ingestion	:	HBsII Lite Reagent	No specific data.
		HBsII Solid Phase Reagent	No specific data.
		HBsII Ancillary Reagent	No specific data.
		HBsII Low Calibrator	No specific data.
		HBsII High Calibrator	No specific data.

Delayed and immediate effects and also chronic effects from short and long term exposure

Short term exposure

Potential immediate effects	:	HBsII Lite Reagent	Not available.
		HBsII Solid Phase Reagent	Not available.
		HBsII Ancillary Reagent	Not available.
		HBsII Low Calibrator	Not available.
		HBsII High Calibrator	Not available.
Potential delayed effects	:	HBsII Lite Reagent	Not available.
		HBsII Solid Phase Reagent	Not available.
		HBsII Ancillary Reagent	Not available.
		HBsII Low Calibrator	Not available.
		HBsII High Calibrator	Not available.

Long term exposure

Section 11. Toxicological information

Potential immediate effects	:	HBsII Lite Reagent	Not available.
		HBsII Solid Phase Reagent	Not available.
		HBsII Ancillary Reagent	Not available.
		HBsII Low Calibrator	Not available.
		HBsII High Calibrator	Not available.
Potential delayed effects	:	HBsII Lite Reagent	Not available.
		HBsII Solid Phase Reagent	Not available.
		HBsII Ancillary Reagent	Not available.
		HBsII Low Calibrator	Not available.
		HBsII High Calibrator	Not available.

Potential chronic health effects

Not available.

Conclusion/Summary	:	Not available.	HBsII Lite Reagent
		Not available.	HBsII Solid Phase Reagent
		Not available.	HBsII Ancillary Reagent
		Not available.	HBsII Low Calibrator
		Not available.	HBsII High Calibrator

General : No known significant effects or critical hazards.

Carcinogenicity : No known significant effects or critical hazards.

Mutagenicity : No known significant effects or critical hazards.

Teratogenicity : No known significant effects or critical hazards.

Developmental effects : No known significant effects or critical hazards.

Fertility effects : No known significant effects or critical hazards.

Numerical measures of toxicity

Acute toxicity estimates

Route	ATE value
HBsII Lite Reagent	
Oral	291262.1 mg/kg
HBsII Solid Phase Reagent	
Oral	103448.3 mg/kg
HBsII Ancillary Reagent	
Oral	291262.1 mg/kg

Interactive effects	:	HBsII Lite Reagent	Not available.
		HBsII Solid Phase Reagent	Not available.
		HBsII Ancillary Reagent	Not available.
		HBsII Low Calibrator	Not available.
		HBsII High Calibrator	Not available.
Other information	:	HBsII Lite Reagent	Not available.
		HBsII Solid Phase Reagent	Not available.
		HBsII Ancillary Reagent	Not available.
		HBsII Low Calibrator	Not available.
		HBsII High Calibrator	Not available.

Section 12. Ecological information

Toxicity

Section 12. Ecological information

Product/ingredient name	Result	Species	Exposure
HBsII Lite Reagent			
sodium azide	Acute EC50 0.348 mg/l Fresh water	Algae - Pseudokirchneriella subcapitata	96 hours
	Acute EC50 4.2 to 6.2 mg/l Fresh water	Daphnia - Daphnia pulex - Larvae	48 hours
	Acute LC50 9000 µg/l Fresh water	Crustaceans - Gammarus lacustris	48 hours
	Acute LC50 0.68 mg/l Fresh water	Fish - Lepomis macrochirus	96 hours
	Chronic NOEC 5600 µg/l Marine water	Algae - Macrocystis pyrifera	96 hours
HBsII Solid Phase Reagent			
sodium azide	Acute EC50 0.348 mg/l Fresh water	Algae - Pseudokirchneriella subcapitata	96 hours
	Acute EC50 4.2 to 6.2 mg/l Fresh water	Daphnia - Daphnia pulex - Larvae	48 hours
	Acute LC50 9000 µg/l Fresh water	Crustaceans - Gammarus lacustris	48 hours
	Acute LC50 0.68 mg/l Fresh water	Fish - Lepomis macrochirus	96 hours
	Chronic NOEC 5600 µg/l Marine water	Algae - Macrocystis pyrifera	96 hours
HBsII Ancillary Reagent			
sodium azide	Acute EC50 0.348 mg/l Fresh water	Algae - Pseudokirchneriella subcapitata	96 hours
	Acute EC50 4.2 to 6.2 mg/l Fresh water	Daphnia - Daphnia pulex - Larvae	48 hours
	Acute LC50 9000 µg/l Fresh water	Crustaceans - Gammarus lacustris	48 hours
	Acute LC50 0.68 mg/l Fresh water	Fish - Lepomis macrochirus	96 hours
	Chronic NOEC 5600 µg/l Marine water	Algae - Macrocystis pyrifera	96 hours
HBsII Low Calibrator			
sodium azide	Acute EC50 0.348 mg/l Fresh water	Algae - Pseudokirchneriella subcapitata	96 hours
	Acute EC50 4.2 to 6.2 mg/l Fresh water	Daphnia - Daphnia pulex - Larvae	48 hours
	Acute LC50 9000 µg/l Fresh water	Crustaceans - Gammarus lacustris	48 hours
	Acute LC50 0.68 mg/l Fresh water	Fish - Lepomis macrochirus	96 hours
	Chronic NOEC 5600 µg/l Marine water	Algae - Macrocystis pyrifera	96 hours
HBsII High Calibrator			
sodium azide	Acute EC50 0.348 mg/l Fresh water	Algae - Pseudokirchneriella subcapitata	96 hours
	Acute EC50 4.2 to 6.2 mg/l Fresh water	Daphnia - Daphnia pulex - Larvae	48 hours
	Acute LC50 9000 µg/l Fresh water	Crustaceans - Gammarus lacustris	48 hours
	Acute LC50 0.68 mg/l Fresh water	Fish - Lepomis macrochirus	96 hours
	Chronic NOEC 5600 µg/l Marine water	Algae - Macrocystis pyrifera	96 hours

Conclusion/Summary :

HBsII Lite Reagent	Not available.
HBsII Solid Phase Reagent	Not available.
HBsII Ancillary Reagent	Not available.
HBsII Low Calibrator	Not available.
HBsII High Calibrator	Not available.

Persistence and degradability

Conclusion/Summary :

HBsII Lite Reagent	Not available.
HBsII Solid Phase Reagent	Not available.
HBsII Ancillary Reagent	Not available.
HBsII Low Calibrator	Not available.
HBsII High Calibrator	Not available.

Bioaccumulative potential

Not available.

Section 12. Ecological information

Mobility in soil

Soil/water partition coefficient (K_{OC})	:	HBsII Lite Reagent	Not available.
		HBsII Solid Phase Reagent	Not available.
		HBsII Ancillary Reagent	Not available.
		HBsII Low Calibrator	Not available.
		HBsII High Calibrator	Not available.
Mobility	:	HBsII Lite Reagent	Not available.
		HBsII Solid Phase Reagent	Not available.
		HBsII Ancillary Reagent	Not available.
		HBsII Low Calibrator	Not available.
		HBsII High Calibrator	Not available.

Other adverse effects : No known significant effects or critical hazards.

Section 13. Disposal considerations

Disposal methods : The generation of waste should be avoided or minimized wherever possible. Disposal of this product, solutions and any by-products should at all times comply with the requirements of environmental protection and waste disposal legislation and any regional local authority requirements. Dispose of surplus and non-recyclable products via a licensed waste disposal contractor. Waste should not be disposed of untreated to the sewer unless fully compliant with the requirements of all authorities with jurisdiction. Waste packaging should be recycled. Incineration or landfill should only be considered when recycling is not feasible. This material and its container must be disposed of in a safe way. Empty containers or liners may retain some product residues. Avoid dispersal of spilled material and runoff and contact with soil, waterways, drains and sewers.

Sodium azide may react with lead or copper plumbing to form highly explosive metal azides.

Section 14. Transport information

DOT Classification

UN number	HBsII Lite Reagent	Not regulated.
	HBsII Solid Phase Reagent	Not regulated.
	HBsII Ancillary Reagent	Not regulated.
	HBsII Low Calibrator	Not regulated.
	HBsII High Calibrator	Not regulated.
UN proper shipping name	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-
Transport hazard class(es)	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-
Packing group	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-

Section 14. Transport information

Environmental hazards	HBsII Lite Reagent	No.
	HBsII Solid Phase Reagent	No.
	HBsII Ancillary Reagent	No.
	HBsII Low Calibrator	No.
	HBsII High Calibrator	No.
Additional information	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-

TDG Classification

UN number	HBsII Lite Reagent	Not regulated.
	HBsII Solid Phase Reagent	Not regulated.
	HBsII Ancillary Reagent	Not regulated.
	HBsII Low Calibrator	Not regulated.
	HBsII High Calibrator	Not regulated.
UN proper shipping name	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-
Transport hazard class(es)	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-
Packing group	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-
Environmental hazards	HBsII Lite Reagent	No.
	HBsII Solid Phase Reagent	No.
	HBsII Ancillary Reagent	No.
	HBsII Low Calibrator	No.
	HBsII High Calibrator	No.
Additional information	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-

ADR/RID

UN number	HBsII Lite Reagent	Not regulated.
	HBsII Solid Phase Reagent	Not regulated.
	HBsII Ancillary Reagent	Not regulated.
	HBsII Low Calibrator	Not regulated.
	HBsII High Calibrator	Not regulated.

Section 14. Transport information

UN proper shipping name	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-
Transport hazard class(es)	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-
Packing group	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-
Environmental hazards	HBsII Lite Reagent	No.
	HBsII Solid Phase Reagent	No.
	HBsII Ancillary Reagent	No.
	HBsII Low Calibrator	No.
	HBsII High Calibrator	No.
Additional information	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-

IMDG

UN number	HBsII Lite Reagent	Not regulated.
	HBsII Solid Phase Reagent	Not regulated.
	HBsII Ancillary Reagent	Not regulated.
	HBsII Low Calibrator	Not regulated.
	HBsII High Calibrator	Not regulated.
UN proper shipping name	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-
Transport hazard class(es)	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-
Packing group	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-

Section 14. Transport information

Environmental hazards	HBsII Lite Reagent	No.
	HBsII Solid Phase Reagent	No.
	HBsII Ancillary Reagent	No.
	HBsII Low Calibrator	No.
	HBsII High Calibrator	No.
Additional information	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-

IATA

UN number	HBsII Lite Reagent	Not regulated.
	HBsII Solid Phase Reagent	Not regulated.
	HBsII Ancillary Reagent	Not regulated.
	HBsII Low Calibrator	Not regulated.
	HBsII High Calibrator	Not regulated.
UN proper shipping name	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-
Transport hazard class(es)	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-
Packing group	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-
Environmental hazards	HBsII Lite Reagent	No.
	HBsII Solid Phase Reagent	No.
	HBsII Ancillary Reagent	No.
	HBsII Low Calibrator	No.
	HBsII High Calibrator	No.
Additional information	HBsII Lite Reagent	-
	HBsII Solid Phase Reagent	-
	HBsII Ancillary Reagent	-
	HBsII Low Calibrator	-
	HBsII High Calibrator	-

Special precautions for user :	HBsII Lite Reagent	**Transport within user's premises:** always transport in closed containers that are upright and secure. Ensure that persons transporting the product know what to do in the event of an accident or spillage.
	HBsII Solid Phase Reagent	**Transport within user's premises:** always transport in closed containers that are upright and secure. Ensure that persons transporting the product know what to do in the event of an accident or spillage.
	HBsII Ancillary Reagent	**Transport within user's premises:** always

Section 14. Transport information

	transport in closed containers that are upright and secure. Ensure that persons transporting the product know what to do in the event of an accident or spillage.
HBsII Low Calibrator	**Transport within user's premises:** always transport in closed containers that are upright and secure. Ensure that persons transporting the product know what to do in the event of an accident or spillage.
HBsII High Calibrator	**Transport within user's premises:** always transport in closed containers that are upright and secure. Ensure that persons transporting the product know what to do in the event of an accident or spillage.

Transport in bulk according to Annex II of MARPOL and the IBC Code

Proper shipping name :

Ship type :

Pollution category :

Section 15. Regulatory information

U.S. Federal regulations : **TSCA 8(a) PAIR**: Poly(oxy-1,2-ethanediyl), α-[4-(1,1,3,3-tetramethylbutyl)phenyl]-ω-hydroxy-

TSCA 8(a) CDR Exempt/Partial exemption: Not determined

United States inventory (TSCA 8b): Not determined.

Clean Water Act (CWA) 311: disodium hydrogenorthophosphate

Clean Air Act Section 112 (b) Hazardous Air Pollutants (HAPs) : Not listed

Clean Air Act Section 602 Class I Substances : Not listed

Clean Air Act Section 602 Class II Substances : Not listed

DEA List I Chemicals (Precursor Chemicals) : Not listed

DEA List II Chemicals (Essential Chemicals) : Not listed

SARA 302/304

Composition/information on ingredients

Name	%	EHS	SARA 302 TPQ (lbs)	SARA 302 TPQ (gallons)	SARA 304 RQ (lbs)	SARA 304 RQ (gallons)
HBsII Lite Reagent sodium azide	0.09	Yes.	500	-	1000	-
HBsII Solid Phase Reagent sodium azide	0.09	Yes.	500	-	1000	-
HBsII Ancillary Reagent sodium azide	0.09	Yes.	500	-	1000	-
HBsII Low Calibrator sodium azide	0.09	Yes.	500	-	1000	-

Section 15. Regulatory information

HBsII High Calibrator sodium azide	0.09	Yes.	500	-	1000	-

SARA 304 RQ : 1111111.1 lbs / 504444.4 kg

SARA 311/312

Classification : Not applicable.

Composition/information on ingredients

Name	%	Fire hazard	Sudden release of pressure	Reactive	Immediate (acute) health hazard	Delayed (chronic) health hazard
HBsII Lite Reagent sodium azide	0.09	No.	No.	No.	Yes.	No.
HBsII Solid Phase Reagent sodium azide	0.09	No.	No.	No.	Yes.	No.
HBsII Ancillary Reagent sodium azide	0.09	No.	No.	No.	Yes.	No.
HBsII Low Calibrator sodium azide	0.09	No.	No.	No.	Yes.	No.
HBsII High Calibrator sodium azide	0.09	No.	No.	No.	Yes.	No.

State regulations

Massachusetts : None of the components are listed.

New York : None of the components are listed.

New Jersey : None of the components are listed.

Pennsylvania : None of the components are listed.

International regulations

Chemical Weapons Convention List Schedule I Chemicals	:	HBsII Lite Reagent	Not listed
		HBsII Solid Phase Reagent	Not listed
		HBsII Ancillary Reagent	Not listed
		HBsII Low Calibrator	Not listed
		HBsII High Calibrator	Not listed
Chemical Weapons Convention List Schedule II Chemicals	:	HBsII Lite Reagent	Not listed
		HBsII Solid Phase Reagent	Not listed
		HBsII Ancillary Reagent	Not listed
		HBsII Low Calibrator	Not listed
		HBsII High Calibrator	Not listed
Chemical Weapons Convention List Schedule III Chemicals	:	HBsII Lite Reagent	Not listed
		HBsII Solid Phase Reagent	Not listed
		HBsII Ancillary Reagent	Not listed
		HBsII Low Calibrator	Not listed
		HBsII High Calibrator	Not listed

Section 16. Other information

History

Date of issue/Date of revision : 5/18/2017

Version : 1.01

Section 16. Other information

Key to abbreviations : ATE = Acute Toxicity Estimate
BCF = Bioconcentration Factor
GHS = Globally Harmonized System of Classification and Labelling of Chemicals
IATA = International Air Transport Association
IBC = Intermediate Bulk Container
IMDG = International Maritime Dangerous Goods
LogPow = logarithm of the octanol/water partition coefficient
MARPOL = International Convention for the Prevention of Pollution From Ships, 1973 as modified by the Protocol of 1978. ("Marpol" = marine pollution)
UN = United Nations

Indicates information that has changed from previously issued version.