Abbott

Release Date: 12/9/16

REF GTIN	Product Name
6K30-10 **00380740010508**	***MULTIGENT Clin Chem Calibrator***

Components:

6K30/10	**MULTIGENT Clin Chem Calibrator**

Abbott Customers:
For additional information, please contact your Abbott Customer Support Center Representative by calling 1-800-527-1869, 1-800-323-9100, or 1-800-235-5396.

Abbott employees:
For additional information relative to the content of the SDSs, please contact your local Safety Representative.

Safety Data Sheet

© Abbott Laboratories Release date 12/09/2016 Last alteration on 12/09/2016

*

1 Identification

· **Product name:** **MULTIGENT Clin Chem Calibrator**

· **ADD List number:** 6K30/10

· **Application of the substance / mixture:** For In Vitro Diagnostic Use

· **Manufacturer / Supplier:**
Abbott Diagnostics
100 Abbott Park Road
Abbott Park, IL 60064-3500

Phone: 1-877-4 ABBOTT

· **Department issuing SDS:** Abbott Diagnostics Environmental Health and Safety

· **Emergency telephone number:**

Contact the CHEMTREC® Emergency Call Center for assistance with transportation or hazardous materials emergencies (24 hours/day, 7 days/week). Refer to Abbott customer number 675805.
- Telephone (800) 424-9300 (toll-free) if you are calling from within the United States, Canada, Puerto Rico and the Virgin Islands.
- Telephone +1 (703) 527-3887, the international and maritime number (collect calls accepted), if you are calling from outside the United States or from a ship at sea.

*

2 Hazard(s) identification

· **Classification of the substance or mixture**

Acute Tox. 4 H302 Harmful if swallowed.

· **Label elements**

· **GHS label elements:** The product is labelled according to the Globally Harmonized System (GHS).

· **Hazard pictograms:**



· **Signal word:** Warning

· **Hazard-determining components of labeling:**
Tetramethylammonium chloride

· **Hazard statements:**
H302 Harmful if swallowed.

· **Precautionary statements:**

P264	Wash hands thoroughly after handling.
P301+P330+P312	IF SWALLOWED: Rinse mouth. Call a POISON CENTER or doctor / physician if you feel unwell.
P501	Dispose of contents / container in accordance with local regulations.

· **Additional information:** Combustible.

· **Routes of Exposure:**

For bloodborne pathogens and potentially infectious materials:
- non-intact skin
- mucous membranes (which includes, but is not limited to, the lining of the nose, mouth and throat)
- parenteral contact (e.g. by injection, puncture)

· **Hazard Overview**
· **Health:** Harmful if swallowed
· **Fire:** Noncombustible
· **Reactivity:** Minimal hazard - Stable, even in a fire. Not reactive with water. Not an oxidizer.

· **Other hazards**
This product contains potentially infectious material. Refer to the US OSHA Bloodborne pathogens standard (29 CFR 1910.1030) for additional relevant information.

*
3 Composition/information on ingredients

· **Chemical characterization:** Mixture of chemical and/or biological substances for in vitro diagnostic use.

· **Hazardous chemical ingredients per U.S. OSHA criteria (29 CFR 1910.1200 Hazard Communication):**		
CAS: 75-57-0	Tetramethylammonium chloride	2.9999%

*
4 First-aid measures

· **After inhalation:** Remove from source of exposure. Seek medical attention and appropriate follow-up.

· **After skin contact:**
Take off any clothing that the product touched. Wash affected area with soap and water. Seek medical attention and appropriate follow-up.

· **After eye contact:**
Rinse open eye(s) cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. Seek medical attention and appropriate follow-up. Wash hands after handling.

· **After swallowing:** Rinse mouth with water. Seek medical attention and appropriate follow-up.

· **Information for Medical Personnel**
This product contains human-sourced and/or potentially infectious material. No known test method can offer complete assurance that products derived from human sources or inactivated microorganisms will not transmit infection. Therefore, all human-sourced material should be considered potentially infectious.

The human-sourced material used in this product has been tested and found to be:
- Nonreactive for HBsAg (hepatitis B surface antigen)
- Nonreactive for anti-HCV (antibodies to hepatitis C virus)
- Nonreactive for anti-HIV-1 (antibodies to human immunodeficiency virus type 1)
- Nonreactive for anti-HIV-2 (antibodies to human immunodeficiency virus type 2)
- Nonreactive for HIV-1 NAT
- Nonreactive for HCV NAT

· **Most important symptoms and effects, both acute and delayed:** None expected

Product name: MULTIGENT Clin Chem Calibrator

· **Medical conditions aggravated by exposure:** None known

5 Fire-fighting measures

· **Suitable extinguishing agents**
Dry chemical, carbon dioxide (CO2), water spray or regular foam.
- Caution: CO2 will displace air in confined spaces and may cause an oxygen-deficient atmosphere.
- For larger fires: There are no unique chemical or reactivity hazards that would impact firefighting decisions related to this product. Use firefighting measures that suit the environment.

· **Special hazards arising from the substance or mixture**
There are no unique chemical or reactivity hazards that would impact firefighting decisions due to the chemicals in this product.

· **Protective equipment**
For large fires, wear appropriate heat- and flame-resistant personal protective equipment and a NFPA/NIOSH approved positive-pressure, self-contained breathing apparatus.

6 Accidental release measures

· **Personal precautions, protective equipment and emergency procedures**
Handle as a potentially infectious material.
Avoid formation of dust.
If material is released or spilled, minimize exposure by using appropriate personal protective equipment as listed in Section 8. Keep unprotected persons away.

· **Environmental precautions** Prevent from entering sewage system, storm drains, surface waters, or soil.

· **Methods and material for containment and cleaning up**
Damp down dust with water spray.
Sweep up material and place into a suitable disposal container.
Vacuum (HEPA filter or equivalent recommended) or wet sweep products that are powders to avoid dust dispersal.

Clean the affected area. Suitable cleaners are:
- warm water and detergent or similar cleansing agent

Apply a suitable disinfectant. Select a disinfectant that is effective against bloodborne infectious agents, as well as other microbial agents that you might expect to be prevalent in your population. A disinfectant that is effective against Mycobacterium tuberculosis is generally effective against all known viruses and non-sporeforming bacteria, and is suitable for most clinical laboratory situations.

NOTE: Commercial disinfectants must be used according to manufacturer directions. Disinfectants are typically hazardous chemicals that react with many chemicals, materials and living tissues. Obtain and review the manufacturer's safety information before using the disinfectant.

Dispose of spilled and contaminated material in accordance with Federal, State, and Local regulations. See Section 13 for information that may impact disposal of materials contaminated with this product.

· **Reference to other sections**
See Section 7 for information on safe handling.
See Section 8 for information on personal protection equipment.

Product name: MULTIGENT Clin Chem Calibrator

See Section 13 for disposal information.

7 Handling and storage

· **Precautions for safe handling:** Handle as a potentially infectious material.

· **Information about protection against explosions and fires:** Combustible - finely divided dust may be explosive.

· **Requirements to be met by storerooms and receptacles:**
Store only in the original container.
Refer to the package insert or product label for additional information on storage conditions for product quality.

· **Information about storage in one common storage facility:** Store in original packaging.

· **Further information about storage conditions:** Protect from heat and direct sunlight.

8 Exposure controls/personal protection

· **Components with Occupational Exposure Limits**
The product does not contain any hazardous ingredients with occupational exposure limits established by OSHA, ACGIH, or NIOSH.

· **General protective and hygienic measures:**
Always maintain good housekeeping and follow general precautionary measures. Do not eat, drink or store food and beverages in areas where chemicals or specimens are used. Wash hands before breaks, after handling reagents and specimens, and at the end of the workshift.

Observe universal precautions and other appropriate biosafety practices for handling potentially infectious material.

· **Breathing equipment:**
Normal use and storage of product - respiratory protection is not necessary if room is well ventilated.

Small-volume spills (e.g. small enough to clean up with a paper towel or small sorbent pad) - respiratory protection should not be necessary if room is well ventilated.

Other unusual conditions (e.g. volume spilled too big to clean up with materials in arm's reach) - Use appropriate NIOSH-approved air-purifying respirator if airborne chemical concentrations may exceed the exposure limit (if any) listed above.

Hazardous Materials Emergencies or Firefighting - use NIOSH/NFPA-approved respiratory protection.

· **Hand protection:**
Wear impervious gloves if hand contact with the material is anticipated. Dispose of contaminated gloves after use in accordance with applicable laws and good laboratory practices.

· **Material of gloves and breakthrough time of the glove material:**
The glove material must be suitable for use in a microbiological laboratory and have a measured breakthrough time of at least 30 minutes, such as those with a Class 2 protection index per EN374 (or equivalent standard applicable in your region). NOTE: This recommendation applies only to the product stated in this Safety Data Sheet. When dissolving in or mixing with other substances, contact the supplier of approved gloves.

· **Eye protection:**
Wear safety glasses or other protective eyewear. If splash potential exists, wear full face shield or goggles.

Product name: MULTIGENT Clin Chem Calibrator

· **Body protection:**
Normal use: protect personal clothing from spatters and small spills. Wear a laboratory coat (or other protective clothing required by your institution).
Larger spills (e.g. that can saturate cloth): wear appropriate water-repellant covering over clothing.

9 Physical and chemical properties

· **General Information**
· **Form:** Solid
· **Color:** Light yellow

· **Odor:** Odorless

· **pH-value at 20 °C (68 °F)** 6.9
Not applicable.
· **Melting point/Melting range:** Not determined
· **Boiling point/Boiling range:** Not determined

· **Flash point** Not applicable
· **Flammability (solid, gaseous)** Not determined.
· **Auto igniting** Product is not self-igniting.

· **Danger of explosion** Product does not present an explosion hazard.
· **Explosion limits**
· **Lower:** Not determined
· **Upper:** Not determined

· **Density at 20 °C (68 °F)** 1.04 g/cm³ (8.679 lbs/gal)
· **Evaporation rate:** Not applicable.

· **Solubility in / Miscibility with**
· **Water:** Soluble
· **Dynamic:** Not applicable.

10 Stability and reactivity

· **Thermal decomposition / conditions to be avoided**
No decomposition if used and stored according to specifications.

· **Possibility of hazardous reactions:** No dangerous reactions known.

· **Conditions to avoid:** No further relevant information available.

· **Incompatible materials:** No further relevant information available.

· **Hazardous decomposition products:** No dangerous decomposition products known.

Product name: MULTIGENT Clin Chem Calibrator

*

11 Toxicological information

· **Acute toxicity**

· **LD50/LC50 values for hazardous ingredients per OSHA criteria:**

· **Ingredients (100% pure substance/s):**		
CAS: 75-57-0 Tetramethylammonium chloride		
Oral	LD50	50 mg/kg (rat) National Technical Information Service. Vol. OTS0536975 Sources cited: Toxicological Data, compiled by the National Institute of Health (NIH), USA, selected and distributed by Technical Database Services (TDS), New York, 2009

· **Primary toxicological effects of the final product:**
· **Skin irritation:** No irritant effect.
· **Eye irritation:** No irritant effect.

· **Sensitization:** No sensitizing effects known.

· **Additional toxicological information:** None

· **Carcinogenic categories**

· **IARC (International Agency for Research on Cancer)**
None of the ingredients is listed.

· **NTP (National Toxicology Program)**
None of the ingredients is listed.

· **OSHA-Ca (Occupational Safety & Health Administration)**
None of the ingredients is listed.

· **Target organs/systems:** Unknown

*

12 Ecological information

· **Aquatic toxicity:**	
CAS: 75-57-0 Tetramethylammonium chloride	
LC50 96 h	462 mg/l (minnow) Source cited: Ecotoxicological Data, compiled by the US Environmental Protection Agency (EPA), selected and distributed by Technical Database Services (TDS), New York, 2009

· **Additional ecological information**

· **General notes:**
Do not allow undiluted product or large quantities of it to reach ground water, water course, or sewage system.

· **Results of PBT and vPvB assessment**

· **PBT:** Not applicable
· **vPvB:** Not applicable

Product name: MULTIGENT Clin Chem Calibrator

13 Disposal considerations

· **Recommendation for disposal of unused product:**
Dispose in accordance with federal, state and local regulations and institutional requirements. The following may be particularly important when identifying appropriate disposal:
- Potentially infectious. See Section 4, Information for Medical Personnel, for more information.
- See Section 6 for information when institutional or regulatory requirements include any sort of treatment of potentially infectious waste.

· **Recommendation for disposal of packaging:**
Non-contaminated packaging may be used for recycling. Refer to applicable local regulations and institutional policies.
For disposal of contaminated packaging, refer to applicable local regulations and institutional policies.

· **Recommended cleansing agent:** Water with cleansing agents, if necessary.

*

14 Transport information

· **DOT, ADN, IMDG, IATA** none

· **UN proper shipping name**
· **DOT, ADR, ADN, IMDG, IATA** none

· **Transport hazard class(es)**

· **DOT, ADR, ADN, IMDG, IATA**
· **Class** none
· **DOT, IMDG, IATA** none

· **Environmental hazards**
· **Marine pollutant:** No

· **Additional information**

· **DOT**
· **Remarks:** Not restricted for transportation.
· **ADR**
· **Remarks:** Not restricted for transportation.
· **IMDG**
· **Remarks:** Not restricted for transportation.
· **IATA**
· **Remarks:** Not restricted for transportation.

*

15 Regulatory information

· **SARA (Superfund Amendments and Reauthorization Act of 1986 - USA):**

· **Section 302/304 (40CFR355.30 / 40CFR355.40):**

The product does not contain listed substances.

Product name: MULTIGENT Clin Chem Calibrator

· **Section 313 (40CFR372.65):**

The product does not contain listed substances.

· **California Proposition 65 (USA):**

· **Chemicals known to cause cancer:**

The product does not contain listed substances.

· **Chemicals known to cause female reproductive toxicity:**

None of the ingredients is listed.

· **Chemicals known to cause male reproductive toxicity:**

None of the ingredients is listed.

· **Chemicals known to cause developmental reproductive toxicity:**

None of the ingredients is listed.

*

16 Other information

The information and recommendations contained herein are based upon information or tests believed to be reliable. Abbott Laboratories does not guarantee the accuracy or completeness of this information or recommendations contained herein, NOR SHALL ANY OF THIS INFORMATION CONSTITUTE A WARRANTY, WHETHER EXPRESSED OR IMPLIED, AS TO THE SAFETY OF THE GOODS, THE MERCHANTABILITY OF THE GOODS, OR THE FITNESS OF THE GOODS FOR A PARTICULAR PURPOSE.

This information is not a substitute for the advice of a health care professional, nor is it a recommendation for any particular course of treatment. It is not intended to supplement, modify or supersede any information (e.g. labeling and package inserts) provided with respect to the medical use of the product. Abbott Laboratories assumes no responsibility for results obtained or for incidental or consequential damages, including lost profits, arising from the use of these data. No warranty against infringement of any patent, copyright or trademark is made or implied.

· **Department issuing SDS**

- Abbott Diagnostics Safety, Health and Environmental Assurance
Department 0571

· **Contact**

- General information about this product:

Abbott Diagnostics
Technical Support
100 Abbott Park Road
Abbott Park, IL 60064-3500

Phone: 1-877-4 ABBOTT

· **Date of preparation / last revision** 12/09/2016 / 13

· **Abbreviations and acronyms:**

RID: Règlement international concernant le transport des marchandises dangereuses par chemin de fer (Regulations Concerning the International Transport of Dangerous Goods by Rail)
ICAO: International Civil Aviation Organisation
ADR: Accord européen sur le transport des marchandises dangereuses par Route (European Agreement concerning the International Carriage of Dangerous Goods by Road)
IMDG: International Maritime Code for Dangerous Goods
DOT: US Department of Transportation
IATA: International Air Transport Association
EINECS: European Inventory of Existing Commercial Chemical Substances

Product name: MULTIGENT Clin Chem Calibrator

ELINCS: European List of Notified Chemical Substances
CAS: Chemical Abstracts Service (Division of the American Chemical Society)
LC50: Lethal concentration, 50 percent
LD50: Lethal dose, 50 percent
PBT: persistent, bioaccumulative and toxic
vPvB: very persistent and very bioaccumulative
OSHA: Occupational Safety & Health
TLV: Threshold Limit Value
PEL: Permissible Exposure Limit
REL: Recommended Exposure Limit
Acute Tox. 4: Acute toxicity Category 4

· *** Sections marked with an asterisk (*) have been altered since the previous version.**